

Appendix M. MAUDE Reports

Table M1. Reports on CGM Devices from the FDA MAUDE Database

Note: Key information is presented in this table, but detailed event descriptions are available at the [FDA MAUDE database](#).

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5161155	10/6/2024 4:00	Injury	DEXCOM, INC.	10/16/2024 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5160563		Death	DEXCOM, INC.	10/3/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		Insufficient Information
MW5160567	9/27/2024 4:00	Injury	DEXCOM, INC.	10/3/2024 4:00	QBJ	DEXCOM G6 SENSOR		Hyperglycemia; Hypoglycemia
MW5160467		Injury	DEXCOM, INC.	10/2/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions
MW5160472		Injury	DEXCOM, INC.	10/2/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5160333		Death	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5160384		Malfunction	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR SYSTEM		Rash
MW5160420		Injury	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM G6		Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5160 191		Injury	DEXCOM, INC.	9/26/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		Hyperglycemia
MW5160 099	9/19/2024 4:00	Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM G6 SENSOR		Insufficient Information
MW5160 103		Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions
MW5160 104		Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions
MW5160 013	9/16/2024 4:00	Injury	DEXCOM, INC.	9/24/2024 4:00	QBJ	DEXCOM SENSOR		Insufficient Information
MW5160 023	9/17/2024 4:00	Injury	DEXCOM, INC.	9/24/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159 708	9/16/2024 4:00	Injury	DEXCOM, INC.	9/17/2024 4:00	QBJ	DEXCOM BLOOD GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5159 446	9/5/2024 4:00	Injury	DEXCOM, INC.	9/9/2024 4:00	QBJ	DEXCOM G6 SENSOR	High Test Results	Insufficient Information
MW5159 455	8/2/2024 4:00	Injury	DEXCOM, INC.	9/9/2024 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5159 311	8/31/2024 4:00	Injury	DEXCOM, INC.	9/6/2024 4:00	QBJ	DEXCOM SENSOR		No Clinical Signs, Symptoms or Conditions
MW5159 349	9/4/2024 4:00	Injury	DEXCOM, INC.	9/6/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5159 056	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5159 057	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5159 058	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5159 081	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6	High Readings; Device Sensing Problem	Hyperglycemia
MW5159 137	8/27/2024 4:00	Malfunction	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6 SENSOR		Insufficient Information
MW5158 710		Injury	DEXCOM, INC.	8/21/2024 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5158 575	8/14/2024 4:00	Injury	DEXCOM, INC.	8/16/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5158 302		Injury	DEXCOM, INC.	8/9/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5158 257	8/5/2024 4:00	Injury	DEXCOM, INC.	8/7/2024 4:00	QBJ	DEXCOM	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5157 864		Injury	DEXCOM, INC.	7/30/2024 4:00	QBJ	DEXCOM		Hypoglycemia
MW5157 737	7/22/2024 4:00	Injury	DEXCOM, INC.	7/26/2024 4:00	QBJ	DEXCOM G6 SENSOR		Discomfort

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5156 911	6/26/2024 4:00	Malfunction	DEXCOM, INC.	7/1/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5156 505	6/17/2024 4:00	Injury	DEXCOM, INC.	6/21/2024 4:00	QBJ	DEXCOM G6		Insufficient Information; No Clinical Signs, Symptoms or Conditions
MW5156 459		Injury	DEXCOM, INC.	6/20/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5156 471		Injury	DEXCOM, INC.	6/20/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5155 988		Injury	DEXCOM, INC.	6/7/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5155 768	6/3/2024 4:00	Malfunction	DEXCOM, INC.	6/4/2024 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5153 878	4/11/2024 4:00	Malfunction	DEXCOM, INC.	4/15/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5153 513	4/3/2024 4:00	Injury	DEXCOM, INC.	4/3/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5153 215	3/23/2024 4:00	Injury	DEXCOM, INC.	3/26/2024 4:00	QBJ	DEXCOM SENSOR	High Readings	No Clinical Signs, Symptoms or Conditions
MW5153 027	3/18/2024 4:00	Injury	DEXCOM, INC.	3/20/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5153005	2/20/2024 5:00	Malfunction	DEXCOM, INC.	3/19/2024 4:00	QBJ	DEXCOM G6 CONTINUEO US GLUCOSE MONITOR SYSTEM		Foreign Body In Patient
MW5152444		Injury	DEXCOM, INC.	3/6/2024 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5151801	2/11/2024 5:00	Injury	DEXCOM, INC.	2/21/2024 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5151825		Injury	DEXCOM, INC.	2/21/2024 5:00	QBJ	DEXCOM	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5151167	2/5/2024 5:00	Injury	DEXCOM, INC.	2/7/2024 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5150660	1/20/2024 5:00	Injury	DEXCOM, INC.	1/23/2024 5:00	QBJ	DEXCOM G6 SENSOR	High Readings	No Clinical Signs, Symptoms or Conditions
MW5150681	1/21/2024 5:00	Malfunction	DEXCOM, INC.	1/23/2024 5:00	QBJ	DEXCOM G6 TRANSMITTER		No Clinical Signs, Symptoms or Conditions
MW5150490		Injury	DEXCOM, INC.	1/17/2024 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5150101		Injury	DEXCOM, INC.	1/9/2024 5:00	QBJ	DEXCOM G6	High Readings; Device Sensing Problem	Hyperglycemia
MW5149414		Injury	DEXCOM, INC.	12/19/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia
MW5148679		Injury	DEXCOM, INC.	12/1/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	No Clinical Signs, Symptoms or Conditions
MW5148559		Injury	DEXCOM, INC.	11/29/2023 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Failure to Sense	Hyperglycemia
MW5148569	11/23/2023 5:00	Injury	DEXCOM, INC.	11/29/2023 5:00	QBJ	DEXCOM G6 CGM SENSOR		Unspecified Infection; Rash
MW5148448		Injury	DEXCOM, INC.	11/22/2023 5:00	QBJ	DEXCOM G6	High Readings; Calibration Problem	Insufficient Information
MW5148455	11/18/2023 5:00	Malfunction	DEXCOM, INC.	11/22/2023 5:00	QBJ	DEXCOM G6 CGM SENSOR		No Clinical Signs, Symptoms or Conditions
MW5148368		Injury	DEXCOM, INC.	11/21/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia
MW5148145		Injury	DEXCOM, INC.	11/14/2023 5:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5148148		Injury	DEXCOM, INC.	11/14/2023 5:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5148 101		Injury	DEXCOM, INC.	11/13/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5148 022		Injury	DEXCOM, INC.	11/9/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia
MW5147 844		Injury	DEXCOM, INC.	11/6/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia
MW5147 801		Injury	DEXCOM, INC.	11/3/2023 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5147 758		Injury	DEXCOM, INC.	11/2/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5147 654		Injury	DEXCOM, INC.	11/1/2023 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5147 085	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR		Insufficient Information
MW5147 086	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR		Insufficient Information
MW5147 087	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR	Defective Device	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5146 930	10/8/2023 4:00	Injury	DEXCOM, INC.	10/12/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia; Hypoglycemia; Diabetic Ketoacidosis
MW5146 931		Injury	DEXCOM, INC.	10/12/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146 871		Injury	DEXCOM, INC.	10/11/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5146 505		Injury	DEXCOM, INC.	10/3/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5146 340		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146 355		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146 362		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146 370		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146 223	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR		No Clinical Signs, Symptoms or Conditions
MW5146 224	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5146 225	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR TRANSMITTER		No Clinical Signs, Symptoms or Conditions
MW5146 203		Injury	DEXCOM, INC.	9/26/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146 144	9/20/2023 4:00	Injury	DEXCOM, INC.	9/25/2023 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5146 150		Injury	DEXCOM, INC.	9/25/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146 110		Injury	DEXCOM, INC.	9/22/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145 980		Injury	DEXCOM, INC.	9/20/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145 922		Injury	DEXCOM, INC.	9/19/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5145 630	9/9/2023 4:00	Injury	DEXCOM, INC.	9/13/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5145 634	9/9/2023 4:00	Injury	DEXCOM, INC.	9/13/2023 4:00	QBJ	DEXCOM		Skin Inflammation/Irritation
MW5145 381		Injury	DEXCOM, INC.	9/7/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145 386		Injury	DEXCOM, INC.	9/7/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145 275		Injury	DEXCOM, INC.	9/6/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5144 951		Injury	DEXCOM, INC.	8/24/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5144 952		Injury	DEXCOM, INC.	8/24/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5144 652		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5144 656		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5144 657		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5122 881	7/25/2023 4:00	Injury	DEXCOM, INC.	8/9/2023 4:00	QBJ	DEXCOM G6		Dehydration; Diarrhea; Fatigue; Hyperglycemia; Vomiting; Diabetic Ketoacidosis
MW5120 629		Injury	DEXCOM, INC.	8/4/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5120 350		Injury	DEXCOM, INC.	7/28/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5120 217		Injury	DEXCOM, INC.	7/26/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5119 983		Injury	DEXCOM, INC.	7/21/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5119 229	7/5/2023 4:00	Injury	DEXCOM, INC.	7/7/2023 4:00	QBJ	DEXCOM G6 SENSOR	Pumping Stopped; Low Readings	Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5118 823	6/21/2023 4:00	Injury	DEXCOM, INC.	6/23/2023 4:00	MDS	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5118 379		Injury	DEXCOM, INC.	6/13/2023 4:00	QBJ	DEXCOM G6	Low Readings; Calibration Problem	No Clinical Signs, Symptoms or Conditions
MW5118 353		Injury	DEXCOM, INC.	6/12/2023 4:00	QBJ	DEXCOM G6	Low Readings	Hyperglycemia
MW5118 227	6/2/2023 4:00	Injury	DEXCOM, INC.	6/6/2023 4:00	MDS	DEXCOM G6 SENSOR	High Readings	No Clinical Signs, Symptoms or Conditions
MW5118 234	6/4/2023 4:00	Malfunction	DEXCOM, INC.	6/6/2023 4:00	QBJ	DEXCOM G6	Loss of or Failure to Bond	No Clinical Signs, Symptoms or Conditions
MW5118 248	6/3/2023 4:00	Injury	DEXCOM, INC.	6/6/2023 4:00	QBJ	DEXCOM G6 SENSOR	Patient Device Interaction Problem	Erythema; Itching Sensation; Pain; Rash; Fluid Discharge; Skin Inflammation/Irritation; Swelling/Edema
MW5118 087		Injury	DEXCOM, INC.	6/1/2023 4:00	MDS	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5118 102		Injury	DEXCOM, INC.	6/1/2023 4:00	QBJ	DEXCOM G6	Chemical Problem; Patient Device Interaction Problem	Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5118033	5/28/2023 4:00	Injury	DEXCOM, INC.	5/31/2023 4:00	QBJ	DEXCOM G6	High Readings	Malaise
MW5117987		Injury	DEXCOM, INC.	5/30/2023 4:00	QBJ	DEXCOM G6/G7	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5117994		Injury	DEXCOM INC.	5/30/2023 4:00	MDS	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Failure to Analyze Signal; Wireless Communication Problem	Hypoglycemia
MW5117996	5/16/2023 4:00	Injury	DEXCOM, INC.	5/30/2023 4:00	QBJ	DEXCOM G6	Failure to Calibrate	Hyperglycemia; Hypoglycemia
MW5117969	5/25/2023 4:00	Injury	DEXCOM, INC.	5/26/2023 4:00	MDS	DEXCOM G6 SENSOR	High Readings	Hypoglycemia
MW5117806	5/22/2023 4:00	Injury	DEXCOM, INC.	5/24/2023 4:00	MDS	DEXCOM G6 SENSOR	Product Quality Problem	Insufficient Information
MW5117807	5/22/2023 4:00	Malfunction	DEXCOM, INC.	5/24/2023 4:00	MDS	DEXCOM G6 SENSOR	Defective Device	No Clinical Signs, Symptoms or Conditions
MW5117795	5/20/2023 4:00	Injury	DEXCOM, INC.	5/23/2023 4:00	MDS	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5117771		Malfunction	DEXCOM, INC.	5/22/2023 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5117 730	5/17/2023 4:00	Malfunction	DEXCOM INC.	5/19/2023 4:00	MDS	DEXCOM G6 SENSOR	Low Readings	No Clinical Signs, Symptoms or Conditions
MW5117 291	4/30/2023 4:00	Injury	DEXCOM, INC.	5/5/2023 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5117 292	4/30/2023 4:00	Injury	DEXCOM, INC.	5/5/2023 4:00	MDS	DEXCOM G6 SENSOR	Entrapment of Device; Incorrect, Inadequate or Imprecise Result or Readings; Connection Problem	Hyperglycemia
MW5117 293	4/30/2023 4:00	Injury	DEXCOM, INC.	5/5/2023 4:00	MDS	DEXCOM G6 SENSOR	Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5117 095	4/21/2023 4:00	Injury	DEXCOM, INC.	4/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Device Displays Incorrect Message	No Clinical Signs, Symptoms or Conditions
MW5117 097	4/21/2021 4:00	Injury	DEXCOM, INC.	4/27/2023 4:00	QBJ	DEXCOM G6	Device Sensing Problem	Hyperglycemia; Hypoglycemia
MW5115 443	2/28/2023 5:00	Malfunction	DEXCOM, INC.	3/2/2023 5:00	QBJ	DEXCOM G6	Application Program Problem; Loss of Data	No Clinical Signs, Symptoms or Conditions
MW5114 604	1/24/2023 5:00	Injury	DEXCOM, INC.	1/31/2023 5:00	QBJ	DEXCOM G6 APP	Defective Device	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5114 583	1/27/2023 5:00	Injury	DEXCOM, INC.	1/30/2023 5:00	QBJ	DEXCOM G6	Delayed Alarm	Hypoglycemia
MW5113 775	12/9/2022 5:00	Injury	DEXCOM, INC.	12/13/2022 5:00	MDS	DEXCOM G6	Patient-Device Incompatibility	Erythema; Hemorrhage/Bleedi ng; Itching Sensation; Rash; Skin Inflammation/ Irritation
MW5112 549	10/6/2022 4:00	Injury	DEXCOM, INC.	10/11/2022 4:00	QBJ	DEXCOM G6 CONTINUO US GLUCOSE METER	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5112 549	10/6/2022 4:00	Injury	DEXCOM, INC.	10/11/2022 4:00	QBJ	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5112 302	9/10/2022 4:00	Injury	DEXCOM, INC.	9/27/2022 4:00	QBJ	DEXCOM G6	High Readings	Crushing Injury; Hypoglycemia
MW5112 031	6/15/2022 4:00	Injury	DEXCOM, INC.	9/13/2022 4:00	QBJ	DEXCOM G6 TRANSMITT ER	Loss of Data	Insufficient Information
MW5111 831	8/24/2022 4:00	Injury	DEXCOM, INC.	8/31/2022 4:00	QBJ	DEXCOM G6 SENSOR AND TRANSMITT ER	Output Problem	Insufficient Information
MW5111 683	8/20/2022 4:00	Injury	DEXCOM, INC.	8/23/2022 4:00	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Hypersensitivity/All ergic reaction; Unspecified Infection; Rash; Skin Inflammation/ Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5111 422	8/5/2022 4:00	Injury	DEXCOM, INC.	8/10/2022 4:00	QBJ	DEXCOM G6 APPLICATION	Product Quality Problem; Application Program Version or Upgrade Problem	Hypoglycemia
MW5111 054	7/16/2022 4:00	Injury	DEXCOM, INC.	7/21/2022 4:00	QBJ	DEXCOM G6	Patient-Device Incompatibility	Erythema; Itching Sensation; Rash
MW5110 452	6/17/2022 4:00	Malfunction	DEXCOM, INC.	6/22/2022 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Calibration Problem	Insufficient Information
MW5110 333	6/11/2022 4:00	Injury	DEXCOM, INC.	6/14/2022 4:00	MDS	DEXCOM G6 SENSORS	Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5110 333	6/11/2022 4:00	Injury	DEXCOM, INC.	6/14/2022 4:00	MDS	DEXCOM G6 SENSORS	Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5110 333	6/11/2022 4:00	Injury	DEXCOM, INC.	6/14/2022 4:00	MDS	DEXCOM G6 SENSORS	Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5110 333	6/11/2022 4:00	Injury	DEXCOM, INC.	6/14/2022 4:00	MDS	DEXCOM G6 SENSORS	Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5109 768	4/6/2022 4:00	Malfunction	DEXCOM, INC.	5/17/2022 4:00	QBJ	DEXCOM G6 SENSOR	Difficult or Delayed Activation	Rash
MW5109 624	4/1/2022 4:00	Injury	DEXCOM, INC.	5/10/2022 4:00	QBJ	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Erythema; Peeling; Blister; Skin Inflammation/ Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5109 507	4/10/2022 4:00	Injury	DEXCOM, INC.	5/3/2022 4:00	QBJ	DEXCOM G6 CGM	Patient-Device Incompatibility	Erythema; Itching Sensation; Local Reaction
MW5109 507	4/10/2022 4:00	Injury	DEXCOM, INC.	5/3/2022 4:00	QBJ	DEXCOM G6 CGM	Patient-Device Incompatibility	Erythema; Itching Sensation; Local Reaction
MW5109 166	4/15/2022 4:00	Injury	DEXCOM, INC.	4/19/2022 4:00	QDK	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Erythema; Rash; Skin Tears
MW5109 180	4/2/2022 4:00	Malfunction	DEXCOM, INC.	4/19/2022 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Markings/Labelling Problem	Hyperglycemia
MW5108 611		Malfunction	DEXCOM, INC.	3/29/2022 4:00	MDS	DEXCOM G6	Low Test Results; Material Integrity Problem	Insufficient Information
MW5107 031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5107 031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITORING SYSTEM TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5107031		Injury	DEXCOM, INC.	1/26/2022 5:00	QBJ	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings; Communication or Transmission Problem	Hypoglycemia; Visual Impairment; Dizziness; Cognitive Changes; Confusion/Disorientation
MW5106846	1/17/2022 5:00	Injury	DEXCOM, INC.	1/20/2022 5:00	QBJ	DEXCOM G6	Product Quality Problem; Patient Device Interaction Problem	No Clinical Signs, Symptoms or Conditions
MW5106846	1/17/2022 5:00	Injury	DEXCOM, INC.	1/20/2022 5:00	QBJ	SENSOR	Product Quality Problem; Patient Device Interaction Problem	No Clinical Signs, Symptoms or Conditions
MW5106533	1/1/2022 5:00	Malfunction	DEXCOM, INC.	1/7/2022 5:00	QBJ	DEXCOM G6 TRANSMITTERS	Defective Device; Device Markings/Labeling Problem	Insufficient Information
MW5106533	1/1/2022 5:00	Malfunction	DEXCOM, INC.	1/7/2022 5:00	QBJ	DEXCOM G6 TRANSMITTERS	Defective Device; Device Markings/Labeling Problem	Insufficient Information
MW5106534	1/5/2022 5:00	Malfunction	DEXCOM, INC.	1/7/2022 5:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Output Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5106 453	12/30/2021 5:00	Injury	DEXCOM, INC.	1/4/2022 5:00	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash; Caustic/Chemical Burns
MW5106 223	12/18/2021 5:00	Injury	DEXCOM, INC.	12/21/2021 5:00	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5106 223	12/18/2021 5:00	Injury	DEXCOM, INC.	12/21/2021 5:00	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5106 223	12/18/2021 5:00	Injury	DEXCOM, INC.	12/21/2021 5:00	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5106 223	12/18/2021 5:00	Injury	DEXCOM, INC.	12/21/2021 5:00	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5106 223	12/18/2021 5:00	Injury	DEXCOM, INC.	12/21/2021 5:00	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5106 187	12/16/2021 5:00	Injury	DEXCOM, INC.	12/20/2021 5:00	QBJ	DEXCOM G6 CGM	Device Markings/Labelling Problem	Insufficient Information
MW5106 187	12/16/2021 5:00	Injury	DEXCOM, INC.	12/20/2021 5:00	QBJ	SENSOR	Device Markings/Labelling Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5106 133	11/1/2020 4:00	Injury	DEXCOM INC	12/16/2021 5:00	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Erythema; Itching Sensation; Rash; Blister
MW5106 145	12/15/2021 5:00	Malfunction	DEXCOM, INC	12/16/2021 5:00	QBJ	DEXCOM G6 CGM 10 DAY ON BODY CONTINUOUS GLUCOSE SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5106 145	12/15/2021 5:00	Malfunction	DEXCOM, INC	12/16/2021 5:00	QBJ	TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5105 309	11/9/2021 5:00	Injury	DEXCOM, INC.	11/12/2021 5:00	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash
MW5105 309	11/9/2021 5:00	Injury	DEXCOM, INC.	11/12/2021 5:00	QBJ	SENSOR	Patient-Device Incompatibility	Rash
MW5105 241	11/2/2021 4:00	Injury	DEXCOM, INC.	11/9/2021 5:00	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Bruise/Contusion; Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Peeling; Skin Discoloration; Skin Inflammation/Irritation
MW5105 247	11/6/2021 4:00	Injury	DEXCOM, INC.	11/9/2021 5:00	QBJ	DEXCOM G6	Patient Device Interaction Problem	Hypoglycemia
MW5105 095	10/1/2021 4:00	Injury	DEXCOM, INC.	11/2/2021 4:00	QBJ	DEXCOM G6 APP FOR ANDROID	Labelling, Instructions for Use or Training Problem; Product Quality Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5104 943	10/15/2021	Injury	DEXCOM, INC.	10/25/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM	Defective Alarm; Computer Software Problem; Device Sensing Problem	Hyperglycemia
MW5104 775	10/15/2021	Malfunction	DEXCOM, INC.	10/19/2021	MDS	DEXCOM G6 GLUCOSE SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5104 775	10/15/2021	Malfunction	DEXCOM, INC.	10/19/2021	MDS	DEXCOM G6 GLUCOSE SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5104 775	10/15/2021	Malfunction	DEXCOM, INC.	10/19/2021	MDS	DEXCOM G6 GLUCOSE SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5104 775	10/15/2021	Malfunction	DEXCOM, INC.	10/19/2021	QBJ	TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5104 732	1/1/2021	Injury	DEXCOM, INC.	10/18/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility; Therapeutic or Diagnostic Output Failure	Scar Tissue; Skin Disorders; Skin Inflammation/Irritation
MW5104 311	9/21/2021	Malfunction	DEXCOM INC.	9/30/2021	QBJ	DEXCOM G6 CGM	Communication or Transmission Problem; Device-Device Incompatibility; Therapeutic or Diagnostic Output Failure	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5104 208	9/23/2021	Injury	DEXCOM, INC.	9/24/2021	MDS	DEXCOM G6 SENSOR	Defective Component	No Clinical Signs, Symptoms or Conditions
MW5104 208	9/23/2021	Injury	DEXCOM, INC.	9/24/2021	QBJ	TRANSMITTER	Defective Component	No Clinical Signs, Symptoms or Conditions
MW5103 803	8/31/2021	Injury	DEXCOM, INC.	9/8/2021	MDS	DEXCOM G6 SENSOR	Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 803	8/31/2021	Injury	DEXCOM, INC.	9/8/2021	MDS	DEXCOM G6 SENSOR	Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 803	8/31/2021	Injury	DEXCOM, INC.	9/8/2021	MDS	DEXCOM G6 SENSOR	Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 716	8/31/2021	Injury	DEXCOM, INC.	9/2/2021	MDS	DEXCOM G6	Defective Component	Emotional Changes
MW5103 674	8/19/2021	Injury	DEXCOM, INC.	9/1/2021	QBJ	DEXCOM G6	Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Defective Device; Calibration Problem	Apnea; Cardiac Arrest; Loss of consciousness; Hypoglycemic Shock
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 457		Malfunction	DEXCOM G6	8/23/2021	QBJ	DEXCOM G6 TRANSMITTER	Use of Device Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5103 185	5/19/2021	Injury	DEXCOM, INC.	8/11/2021	MDS	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5103054	8/4/2021	Malfunction	DEXCOM, INC.	8/5/2021	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5103054	8/4/2021	Malfunction	DEXCOM, INC.	8/5/2021	MDS	SENSORS	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5102993	8/1/2021	Injury	DEXCOM, INC.	8/3/2021	MDS	DEXCOM G6 CONTINUAL GLUCOSE MONITOR	False Alarm; Low Readings; Therapeutic or Diagnostic Output Failure	Hyperglycemia
MW5102502	7/8/2021	Injury	DEXCOM, INC.	7/12/2021	MDS	DEXCOM G6 SENSOR	Defective Component; Separation Failure; Physical Resistance/Sticking	Insufficient Information
MW5102052	6/19/2021	Malfunction	DEXCOM, INC.	6/22/2021	QBJ	DEXCOM G6 TRANSMITTER	Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5102052	6/19/2021	Malfunction	DEXCOM, INC.	6/22/2021	QBJ	DEXCOM G6 TRANSMITTER	Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5102052	6/19/2021	Malfunction	DEXCOM, INC.	6/22/2021	QBJ	DEXCOM G6 TRANSMITTER	Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5102052	6/19/2021	Malfunction	DEXCOM, INC.	6/22/2021	MDS	SENSORS	Expiration Date Error; Device Markings/Labelling Problem; Therapeutic or Diagnostic Output Failure	Insufficient Information
MW5102054	6/22/2021	Injury	DEXCOM, INC.	6/22/2021	QBJ	DEXCOM CGM	Retraction Problem	Insufficient Information
MW5102054	6/22/2021	Injury	DEXCOM, INC.	6/22/2021	QBJ	DEXCOM CGM	Retraction Problem	Insufficient Information
MW5101707		Malfunction	DEXCOM, INC.	6/3/2021	QBJ	DEXCOM G6	Defective Device	No Clinical Signs, Symptoms or Conditions
MW5101707		Malfunction	DEXCOM, INC.	6/3/2021	MDS	SENSOR	Defective Device	No Clinical Signs, Symptoms or Conditions
MW5101606	5/25/2021	Injury	DEXCOM, INC.	5/27/2021	QBJ	DEXCOM G6	Device Ingredient or Reagent Problem	Partial thickness (Second Degree) Burn
MW5101606	5/25/2021	Injury	DEXCOM, INC.	5/27/2021	MDS	DEXCOM G6	Device Ingredient or Reagent Problem	Partial thickness (Second Degree) Burn

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5101 373	5/12/2021	Malfunction	DEXCOM INC.	5/14/2021	QBJ	DEXCOM G6 SENSOR INSERTER	Activation, Positioning or Separation Problem	Insufficient Information
MW5101 154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information
MW5101 154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	MDS	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information
MW5101 154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	MDS	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information
MW5101 154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	MDS	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information
MW5101 154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	MDS	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5101154	4/30/2021	Malfunction	DEXCOM, INC.	5/4/2021	MDS	SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Insufficient Information
MW5101103	4/28/2021	Malfunction	DEXCOM, INC.	4/30/2021	QBJ	DEXCOM G6 TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings; Defective Device	Hypoglycemia
MW5101103	4/28/2021	Malfunction	DEXCOM, INC.	4/30/2021	QBJ	DEXCOM G6 TRANSMITTER	Incorrect, Inadequate or Imprecise Result or Readings; Defective Device	Hypoglycemia
MW5101021	8/24/2020	Injury	DEXCOM, INC.	4/27/2021	QBJ	DEXCOM G6 CGM	Detachment of Device or Device Component	Foreign Body In Patient
MW5101021	8/24/2020	Injury	DEXCOM, INC.	4/27/2021	MDS	SENSOR	Detachment of Device or Device Component	Foreign Body In Patient
MW5101049	4/1/2020	Injury	DEXCOM, INC.	4/27/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Unspecified Infection
MW5100898	4/11/2021	Injury	DEXCOM, INC.	4/20/2021	QBJ	DEXCOM G6	Use of Device Problem; Battery Problem; Complete Loss of Power	Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5100 853	4/14/2021	Malfunction	DEXCOM, INC.	4/19/2021	MDS	DEXCOM G6 SENSOR	Defective Device; Data Problem	Insufficient Information
MW5100 853	4/14/2021	Malfunction	DEXCOM, INC.	4/19/2021	MDS	DEXCOM G6 SENSOR	Defective Device; Data Problem	Insufficient Information
MW5100 853	4/14/2021	Malfunction	DEXCOM, INC.	4/19/2021	MDS	DEXCOM G6 SENSOR	Defective Device; Data Problem	Insufficient Information
MW5100 853	4/14/2021	Malfunction	DEXCOM, INC.	4/19/2021	MDS	DEXCOM G6 SENSOR	Defective Device; Data Problem	Insufficient Information
MW5100 696	4/7/2021	Injury	DEXCOM, INC	4/9/2021	MDS	DEXCOM G6 CGM SENSOR	Patient Device Interaction Problem	Hypersensitivity/Allergic reaction; Rash
MW5100 649	4/2/2021	Injury	DEXCOM INC	4/7/2021	QBJ	DEXCOM G6	Use of Device Problem; Improper or Incorrect Procedure or Method; Separation Failure	Insufficient Information
MW5100 046	1/1/2021	Injury	DEXCOM, INC.	3/16/2021	QBJ	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5100007	2/19/2021	Injury	DEXCOM, INC.	3/15/2021	MDS	DEXCOM G6 SENSOR	Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell	Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation
MW5100007	2/19/2021	Injury	DEXCOM, INC.	3/15/2021	MDS	DEXCOM G6 SENSOR	Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell	Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation
MW5100007	2/19/2021	Injury	DEXCOM, INC.	3/15/2021	MDS	DEXCOM G6 SENSOR	Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell	Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5100007	2/19/2021	Injury	DEXCOM, INC.	3/15/2021	MDS	DEXCOM G6 SENSOR	Loss of or Failure to Bond; Incorrect, Inadequate or Imprecise Result or Readings; Patient-Device Incompatibility; Device Fell	Itching Sensation; Pain; Rash; Scar Tissue; Skin Inflammation/Irritation
MW5100008	3/1/2021	Injury	DEXCOM, INC.	3/15/2021	MDS	DEXCOM G6	Patient-Device Incompatibility	Erythema; Itching Sensation; Superficial (First Degree) Burn
MW5099926	3/8/2021	Injury	DEXCOM, INC.	3/10/2021	QBJ	DEXCOM G6 CGM	Entrapment of Device; Separation Failure; Physical Resistance/Sticking	Pain
MW5099926	3/8/2021	Injury	DEXCOM, INC.	3/10/2021	MDS	SENSOR	Entrapment of Device; Separation Failure; Physical Resistance/Sticking	Pain
MW5099769	3/1/2021	Injury	DEXCOM, INC.	3/3/2021	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Rash
MW5099774	2/15/2021	Injury	DEXCOM, INC.	3/3/2021	QBJ	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Erythema; Itching Sensation; Swelling

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 774	2/15/2021	Injury	DEXCOM, INC.	3/3/2021	MDS	SENSOR	Adverse Event Without Identified Device or Use Problem	Erythema; Itching Sensation; Swelling
MW5099 749	2/24/2021	Injury	DEXCOM INC.	3/2/2021	QBJ	DEXCOM G6	Patient-Device Incompatibility	Reaction
MW5099 721	2/25/2021	Injury	DEXCOM, INC.	3/1/2021	MDS	DEXCOM G6 CGM	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Burning Sensation; Blister
MW5099 627	2/1/2021	Injury	DEXCOM INC.	2/24/2021	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation; Swelling
MW5099 627	2/1/2021	Injury	DEXCOM INC.	2/24/2021	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation; Swelling
MW5099 627	2/1/2021	Injury	DEXCOM INC.	2/24/2021	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation; Swelling
MW5099 627	2/1/2021	Injury	DEXCOM INC.	2/24/2021	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation; Swelling
MW5099 574	11/2/2022 4:00	Injury	DEXCOM, INC.	2/23/2021 5:00	QBJ	DEXCOM G6 CONTINUOUS US GLUCOSE MONITOR	Inaccurate Delivery; Unexpected Shutdown	Hypoglycemia; Overdose; Diabetic Ketoacidosis
MW5099 574	11/2/2022 4:00	Injury	DEXCOM, INC.	2/23/2021 5:00	QBJ	DEXCOM G6 CONTINUOUS US GLUCOSE MONITOR	Inaccurate Delivery; Unexpected Shutdown	Hypoglycemia; Overdose; Diabetic Ketoacidosis

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 574	11/2/2022 4:00	Injury	DEXCOM, INC.	2/23/2021 5:00	QBJ	DEXCOM SENSOR	Inaccurate Delivery; Unexpected Shutdown	Hypoglycemia; Overdose; Diabetic Ketoacidosis
MW5099 574	11/2/2022 4:00	Injury	DEXCOM, INC.	2/23/2021 5:00	QBJ	DEXCOM SENSOR	Inaccurate Delivery; Unexpected Shutdown	Hypoglycemia; Overdose; Diabetic Ketoacidosis
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 578	2/21/2021	Injury	DEXCOM, INC.	2/23/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE METER	Device Displays Incorrect Message; Communication or Transmission Problem	Hypoglycemia
MW5099 534	2/17/2021	Malfunction	DEXCOM, INC.	2/22/2021	QBJ	DEXCOM G6 CGM	Separation Failure	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 497	2/16/2021	Malfunction	DEXCOM, INC.	2/18/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM SENSOR	Activation Failure	Rash; Local Reaction
MW5099 497	2/16/2021	Malfunction	DEXCOM, INC.	2/18/2021	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM SENSOR	Activation Failure	Rash; Local Reaction
MW5099 459		Injury	DEXCOM, INC.	2/17/2021	MDS	DEXCOM G6	Difficult or Delayed Separation; Erratic Results	Rash
MW5099 311	2/6/2021	Injury	DEXCOM, INC.	2/9/2021	MDS	DEXCOM G6 CGM	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia; Diabetic Ketoacidosis
MW5099 311	2/6/2021	Injury	DEXCOM, INC.	2/9/2021	QBJ	DEXCOM G6 CGM	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia; Diabetic Ketoacidosis
MW5099 325	2/8/2021	Malfunction	DEXCOM, INC.	2/9/2021	MDS	DEXCOM G-6 SENSOR	Adverse Event Without Identified Device or Use Problem	Erythema; Itching Sensation; Rash
MW5099 325	2/8/2021	Malfunction	DEXCOM, INC.	2/9/2021	MDS	DEXCOM G-6 SENSOR	Adverse Event Without Identified Device or Use Problem	Erythema; Itching Sensation; Rash

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 325	2/8/2021	Malfunction	DEXCOM, INC.	2/9/2021	MDS	DEXCOM G-6 SENSOR	Adverse Event Without Identified Device or Use Problem	Erythema; Itching Sensation; Rash
MW5099 210	1/31/2021	Injury	DEXCOM, INC.	2/3/2021	MDS	DEXCOM G6	Difficult to Remove	Blister; Skin Inflammation/ Irritation
MW5099 159	12/16/2020	Injury	DEXCOM, INC.	2/2/2021	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash
MW5099 159	12/16/2020	Injury	DEXCOM, INC.	2/2/2021	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash
MW5099 160	10/12/2020	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Patient-Device Incompatibility	Rash
MW5099 160	10/12/2020	Injury	DEXCOM, INC.	2/2/2021	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash
MW5099 165	1/15/2021	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Patient-Device Incompatibility	Unspecified Infection; Burn, Thermal
MW5099 166	12/30/2020	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash
MW5099 166	12/30/2020	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash
MW5099 166	12/30/2020	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5099 170	7/17/2020	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Erythema; Inflammation; Itching Sensation; Skin Irritation
MW5099 171	1/25/2021	Injury	DEXCOM, INC.	2/2/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Rash
MW5099 114	1/1/2021	Injury	DEXCOM, INC.	1/29/2021	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Pain; Rash
MW5099 021	1/23/2021	Malfunction	DEXCOM, INC.	1/26/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOMETER	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia
MW5098 967	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation
MW5098 967	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation
MW5098 967	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation
MW5098 967	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50989 67	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation
MW50989 67	11/26/2020	Injury	DEXCOM INC	1/22/2021	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Itching Sensation; Rash; Skin Irritation
MW50989 15	1/15/2021	Injury	DEXCOM INC	1/20/2021	MDS	DEXCOM G6	Therapeutic or Diagnostic Output Failure	Erythema
MW50989 15	1/15/2021	Injury	DEXCOM INC	1/20/2021	MDS	DEXCOM G6	Therapeutic or Diagnostic Output Failure	Erythema
MW50989 15	1/15/2021	Injury	DEXCOM INC	1/20/2021	MDS	DEXCOM G6	Therapeutic or Diagnostic Output Failure	Erythema
MW50989 16	11/29/2020	Injury	DEXCOM INC.	1/20/2021	QBJ	DEXCOM G6		
MW50988 61	1/12/2021	Malfunction	DEXCOM INC.	1/15/2021	QBJ	DEXCOM G6	Retraction Problem; Physical Resistance/Sticking	Pain

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50988 73	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information
MW50988 73	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information
MW50988 73	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5098873	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information
MW5098873	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information
MW5098873	1/13/2021	Malfunction	DEXCOM INC	1/15/2021	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings; Activation, Positioning or Separation Problem	Insufficient Information
MW5098827	1/4/2021	Injury	DEXCOM, INC.	1/13/2021	MDS	DEXCOM G6 SENSOR	Mechanical Problem	Hemorrhage/Bleeding; Pain

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5098719	5/1/2020	Injury	DEXCOM INC	1/6/2021	MDS	DEXCOM G6 CGM	Patient-Device Incompatibility	Itching Sensation; Pain; Scarring; Reaction; Caustic/Chemical Burns
MW5098608	9/9/2020	Injury	DEXCOM, INC.	12/31/2020	MDS	DEXCOM G6 CGM	Patient-Device Incompatibility	Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Caustic/Chemical Burns
MW5098608	9/9/2020	Injury	DEXCOM, INC.	12/31/2020	MDS	DEXCOM G6 CGM	Patient-Device Incompatibility	Purulent Discharge; Hemorrhage/Bleeding; Itching Sensation; Caustic/Chemical Burns
MW5098614		Injury	DEXCOM, INC.	12/31/2020	QBJ	DEXCOM G6	Application Program Problem; Device-Device Incompatibility	Insufficient Information
MW5098567	12/25/2020	Injury	DEXCOM, INC.	12/29/2020	MDS	DEXCOM G6 SENSOR	Separation Failure	Hyperglycemia
MW5098567	12/25/2020	Injury	DEXCOM, INC.	12/29/2020	MDS	DEXCOM G6 SENSOR	Separation Failure	Hyperglycemia
MW5098457	12/5/2020	Injury	DEXCOM	12/18/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Retraction Problem; Separation Failure	Pain

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5098396	12/12/2020	Malfunction	DEXCOM INC.	12/15/2020	QBJ	DEXCOM G6	No Audible Alarm	Insufficient Information
MW5098407	5/3/2020	Injury	DEXCOM INC.	12/15/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	Hemorrhage/Bleeding; Itching Sensation; Rash; Blister
MW5098407	5/3/2020	Injury	DEXCOM INC.	12/15/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	Hemorrhage/Bleeding; Itching Sensation; Rash; Blister
MW5098407	5/3/2020	Injury	DEXCOM INC.	12/15/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	Hemorrhage/Bleeding; Itching Sensation; Rash; Blister
MW5098407	5/3/2020	Injury	DEXCOM INC.	12/15/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	Hemorrhage/Bleeding; Itching Sensation; Rash; Blister
MW5098407	5/3/2020	Injury	DEXCOM INC.	12/15/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	Hemorrhage/Bleeding; Itching Sensation; Rash; Blister

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50983 30	11/22/2020	Injury	DEXCOM, INC.	12/11/2020	MDS	DEXCOM G6 CONTINUOU S GLUCOSE MONITORIN G (CGM) SYSTEM	Patient-Device Incompatibility	Local Reaction; Scarring; Burning Sensation
MW50983 30	11/22/2020	Injury	DEXCOM, INC.	12/11/2020	MDS	DEXCOM G6 CONTINUOU S GLUCOSE MONITORIN G (CGM) SYSTEM	Patient-Device Incompatibility	Local Reaction; Scarring; Burning Sensation
MW50981 36	11/23/2020	Injury	DEXCOM, INC.	11/27/2020	QBJ	DEXCOM G6 SENSOR (FOR CONTINUOU S GLUCOSE MONITOR)	Difficult to Remove	Bruise/Contusion; Irritation; Swelling; Blood Loss
MW50977 87	6/1/2020	Injury	DEXCOM, INC.	11/10/2020	MDS	DEXCOM G6	Patient-Device Incompatibility	Erythema; Itching Sensation; Rash
MW50974 91	10/21/2020	Malfunction	DEXCOM, INC.	10/27/2020	MDS	DEXCOM G6 SENSOR	Detachment of Device or Device Component	No Clinical Signs, Symptoms or Conditions
MW50974 97	10/25/2020	Injury	DEXCOM, INC.	10/27/2020	QBJ	DEXCOM G6 CONTINUOU S GLUCOSE MONITOR	Adverse Event Without Identified Device or Use Problem	Skin Erosion; Skin Irritation
MW50974 97	10/25/2020	Injury	DEXCOM, INC.	10/27/2020	MDS	SENSOR	Adverse Event Without Identified Device or Use Problem	Skin Erosion; Skin Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50974 97	10/25/2020	Injury	DEXCOM, INC.	10/27/2020	MDS	SENSOR	Adverse Event Without Identified Device or Use Problem	Skin Erosion; Skin Irritation
MW50974 69	10/18/2020	Injury	DEXCOM, INC.	10/26/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Adverse Event Without Identified Device or Use Problem	Reaction
MW50974 69	10/18/2020	Injury	DEXCOM, INC.	10/26/2020	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Adverse Event Without Identified Device or Use Problem	Reaction
MW50974 29		Malfunction	DEXCOM, INC.	10/23/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITORING SYSTEM (CGM)	Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW50973 79	5/1/2020	Injury	DEXCOM, INC.	10/21/2020	QBJ	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Rash; Scarring; Blister
MW50973 79	5/1/2020	Injury	DEXCOM, INC.	10/21/2020	MDS	SENSORS	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Rash; Scarring; Blister
MW50972 43	10/12/2020	Injury	DEXCOM, INC.	10/14/2020	MDS	DEXCOM G6 GLUCOSE SENSOR	Adverse Event Without Identified Device or Use Problem	Skin Irritation; Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50971 24	10/4/2020	Injury	DEXCOM, INC.	10/6/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Premature Discharge of Battery; Incorrect, Inadequate or Imprecise Result or Readings	Diabetic Ketoacidosis
MW50971 30	10/2/2020	Malfunction	DEXCOM, INC.	10/6/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Unable to Obtain Readings; Loss of Data; Inadequate User Interface	No Clinical Signs, Symptoms or Conditions
MW50971 30	10/2/2020	Malfunction	DEXCOM, INC.	10/6/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Unable to Obtain Readings; Loss of Data; Inadequate User Interface	No Clinical Signs, Symptoms or Conditions
MW50970 30	9/29/2020	Injury	DEXCOM, INC.	10/1/2020	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Erythema; Rash
MW50969 25	9/28/2020	Malfunction	DEXCOM, INC.	9/28/2020	QBJ	DEXCOM G6 CGM	Break	No Clinical Signs, Symptoms or Conditions
MW50969 25	9/28/2020	Malfunction	DEXCOM, INC.	9/28/2020	---	TRANSMITTER WITH DEXCOM CGM	Break	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50968 11	9/16/2020	Injury	DEXCOM, INC.	9/22/2020	MDS	DEXCOM CGM G6	Retraction Problem; Activation, Positioning or Separation Problem; Material Twisted/Bent	Insufficient Information
MW50968 11	9/16/2020	Injury	DEXCOM, INC.	9/22/2020	MDS	DEXCOM CGM G6	Retraction Problem; Activation, Positioning or Separation Problem; Material Twisted/Bent	Insufficient Information
MW50967 39	8/25/2020	Injury	DEXCOM, INC.	9/17/2020	QDK	DEXCOM G6 SENSOR	Output Problem; Physical Resistance/Sticki ng	Hemorrhage/Bleedi ng; Skin Discoloration
MW50967 39	8/25/2020	Injury	DEXCOM, INC.	9/17/2020	QDK	DEXCOM G6 SENSOR	Output Problem; Physical Resistance/Sticki ng	Hemorrhage/Bleedi ng; Skin Discoloration
MW50964 99	9/4/2020	Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM	Product Quality Problem	Scarring; Fluid Discharge
MW50965 08	9/5/2020	Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM G6 SENSOR	Break; Product Quality Problem; Defective Device	Foreign Body In Patient

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50965 10		Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM G6 TRANSMITTE R	Product Quality Problem	Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge
MW50965 10		Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM G6 TRANSMITTE R	Product Quality Problem	Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge
MW50965 10		Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM G6 TRANSMITTE R	Product Quality Problem	Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge
MW50965 10		Injury	DEXCOM, INC.	9/9/2020	MDS	DEXCOM G6 TRANSMITTE R	Product Quality Problem	Hypersensitivity/All ergic reaction; Itching Sensation; Rash; Fluid Discharge
MW50965 11	9/1/2020	Malfunction	DEXCOM, INC.	9/9/2020	QBJ	DEXCOM G6	Loss of or Failure to Bond; Device Fell	No Clinical Signs, Symptoms or Conditions
MW50965 11	9/1/2020	Malfunction	DEXCOM, INC.	9/9/2020	MDS	SENSOR	Loss of or Failure to Bond; Device Fell	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50964 24	9/1/2020	Injury	DEXCOM, INC.	9/3/2020	MDS	DEXCOM G6 TRANSMITTE R ADHESIVE	Patient-Device Incompatibility	Itching Sensation; Rash; Burning Sensation
MW50964 32	8/25/2020	Malfunction	DEXCOM, INC.	9/3/2020	QII	DEXCOM G6 PRO	Battery Problem: Low Impedance	No Known Impact Or Consequence To Patient
MW50964 39	8/30/2020	Injury	DEXCOM, INC.	9/3/2020	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia
MW50964 40	8/18/2020	Malfunction	DEXCOM, INC.	9/3/2020	MDS	DEXCOM G6	Product Quality Problem	Rash; Skin Irritation
MW50964 40	8/18/2020	Malfunction	DEXCOM, INC.	9/3/2020	MDS	DEXCOM G6	Product Quality Problem	Rash; Skin Irritation
MW50963 72	8/29/2020	Malfunction	DEXCOM, INC.	9/1/2020	QBJ	DEXCOM G6	Retraction Problem	Hemorrhage/Bleedi ng
MW50963 16	1/15/2020	Injury	DEXCOM, INC.	8/27/2020	MDS	DEXCOMG6	Patient-Device Incompatibility	Unspecified Infection; Rash; Skin Irritation
MW50963 16	1/15/2020	Injury	DEXCOM, INC.	8/27/2020	MDS	SENSOR	Patient-Device Incompatibility	Unspecified Infection; Rash; Skin Irritation
MW50962 62	8/22/2020	Malfunction	DEXCOM, INC.	8/25/2020	QBJ	DEXCOM G6	Communication or Transmission Problem	No Known Impact Or Consequence To Patient
MW50962 39	6/21/2020	Injury	DEXCOM, INC.	8/24/2020	QBJ	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50961 26	8/15/2020	Injury	DEXCOM, INC.	8/18/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Hypersensitivity/All ergic reaction; Unspecified Infection; Itching Sensation; Skin Irritation; Fluid Discharge
MW50961 26	8/15/2020	Injury	DEXCOM, INC.	8/18/2020	MDS	SENSOR	Patient-Device Incompatibility	Hypersensitivity/All ergic reaction; Unspecified Infection; Itching Sensation; Skin Irritation; Fluid Discharge
MW50960 94	8/12/2020	Injury	DEXCOM, INC.	8/17/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Abrasion; Itching Sensation; Swelling; Discharge; Impaired Healing; Caustic/Chemical Burns
MW50960 27	8/7/2020	Injury	DEXCOM, INC.	8/11/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Rash
MW50959 92	5/17/2020	Injury	DEXCOM, INC.	8/10/2020	MDS	DEXCOM G6 CGM	Patient-Device Incompatibility	Local Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50959 93	2/1/2020	Injury	DEXCOM, INC.	8/10/2020	MDS	DEXCOM G6 SENSORS	Patient-Device Incompatibility; Detachment of Device or Device Component	Unspecified Infection; Itching Sensation; Peeling; Skin Irritation; Discharge; Discomfort
MW50959 68	7/14/2020	Injury	DEXCOM, INC.	8/7/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Unspecified Infection; Local Reaction
MW50959 68	7/14/2020	Injury	DEXCOM, INC.	8/7/2020	MDS	DEXCOM G6	Patient-Device Incompatibility	Unspecified Infection; Local Reaction
MW50958 03	7/27/2020	Injury	DEXCOM, INC.	7/29/2020	QBJ	DEXCOM SENSOR G6	Break; Entrapment of Device	Device Embedded In Tissue or Plaque
MW50957 74	7/24/2020	Injury	DEXCOM, INC.	7/28/2020	MDS	DEXCOM G6	Patient-Device Incompatibility	Itching Sensation; Rash; Discharge; Discomfort
MW50957 49	5/27/2020	Injury	DEXCOM	7/27/2020	MDS	DEXCOMG G6 SENSOR	Patient-Device Incompatibility	Rash; Scarring
MW50955 28	6/30/2020	Injury	DEXCOM, INC.	7/14/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Purulent Discharge; Rash; Caustic/Chemical Burns
MW50955 28	6/30/2020	Injury	DEXCOM, INC.	7/14/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Purulent Discharge; Rash; Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50955 28	6/30/2020	Injury	DEXCOM, INC.	7/14/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Purulent Discharge; Rash; Caustic/Chemical Burns
MW50955 13	6/1/2020	Injury	DEXCOM, INC.	7/13/2020	QBJ	DEXCOM SENSOR	Patient-Device Incompatibility	Burn(s); Itching Sensation; Skin Irritation
MW50954 33	7/4/2020	Injury	DEXCOM, INC.	7/8/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Skin Discoloration
MW50954 43	6/29/2020	Injury	DEXCOM, INC.	7/8/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Inflammation
MW50954 43	6/29/2020	Injury	DEXCOM, INC.	7/8/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Inflammation
MW50954 43	6/29/2020	Injury	UNK	7/8/2020	FRO	DRESSING, WOUND, DRUG	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Inflammation
MW50954 06	7/2/2020	Injury	DEXCOM, INC.	7/7/2020	MDS	DEXCOM G6 SENSOR	Physical Resistance/Sticking	Needle Stick/Puncture
MW50954 06	7/2/2020	Injury	DEXCOM, INC.	7/7/2020	MDS	DEXCOM G6 SENSOR	Physical Resistance/Sticking	Needle Stick/Puncture
MW50954 18	5/18/2020	Injury	DEXCOM, INC.	7/7/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash; Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5095418	5/18/2020	Injury	DEXCOM, INC.	7/7/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash; Reaction
MW5095388	7/1/2020	Injury	DEXCOM, INC.	7/6/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash; Blood Loss; Fluid Discharge
MW5095352	5/15/2020	Injury	DEXCOM, INC.	7/2/2020	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Skin Irritation; No Code Available
MW5095352	5/15/2020	Injury	DEXCOM, INC.	7/2/2020	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Skin Irritation; No Code Available
MW5095352	5/15/2020	Injury	DEXCOM, INC.	7/2/2020	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Skin Irritation; No Code Available
MW5095352	5/15/2020	Injury	DEXCOM, INC.	7/2/2020	MDS	DEXCOM G6	Adverse Event Without Identified Device or Use Problem	Skin Irritation; No Code Available
MW5095334	6/25/2020	Injury	DEXCOM, INC.	7/1/2020	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Scarring; Skin Irritation
MW5095334	6/25/2020	Injury	DEXCOM, INC.	7/1/2020	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Scarring; Skin Irritation

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50953 34	6/25/2020	Injury	DEXCOM, INC.	7/1/2020	MDS	DEXCOM G6 CONTINUOU S GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Scarring; Skin Irritation
MW50953 34	6/25/2020	Injury	DEXCOM, INC.	7/1/2020	MDS	DEXCOM G6 CONTINUOU S GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Scarring; Skin Irritation
MW50953 18	5/1/2020	Injury	DEXCOM, INC.	6/30/2020	QBJ	DEXCOM G6 CONTINUOU S GLUCOSE MONITORIN G SYSTEM	Patient-Device Incompatibility	Reaction; Caustic/Chemical Burns
MW50953 18	5/1/2020	Injury	DEXCOM, INC.	6/30/2020	MDS	SENSOR	Patient-Device Incompatibility	Reaction; Caustic/Chemical Burns
MW50952 43	4/20/2020	Injury	DEXCOM, INC.	6/25/2020	MDS	DEXCOM G-6	Patient-Device Incompatibility	Erythema; Itching Sensation
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction
MW50951 66	6/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSORS	Product Quality Problem	Erythema; Irritation; Itching Sensation; Swelling; Reaction
MW50951 70	5/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Purulent Discharge; Erythema; Inflammation; Itching Sensation; Scarring; Skin Irritation; Caustic/Chemical Burns

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5095170	5/12/2020	Injury	DEXCOM, INC.	6/23/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Purulent Discharge; Erythema; Inflammation; Itching Sensation; Scarring; Skin Irritation; Caustic/Chemical Burns
MW5095173	6/1/2020	Injury	DEXCOM, INC.	6/23/2020	QBJ	DEXCOM G6	Insufficient Information	No Known Impact Or Consequence To Patient
MW5095179	6/19/2020	Malfunction	DEXCOM, INC.	6/23/2020	QBJ	DEXCOM G6	Mechanical Problem	No Known Impact Or Consequence To Patient
MW5095185	6/16/2020	Injury	DEXCOM, INC	6/23/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Skin Irritation
MW5095185	6/16/2020	Injury	DEXCOM, INC	6/23/2020	MDS	SENSOR	Patient-Device Incompatibility	Skin Irritation
MW5095134	6/10/2020	Injury	DEXCOM, INC.	6/19/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation
MW5095134	6/10/2020	Injury	DEXCOM, INC.	6/19/2020	MDS	SENSORS	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation
MW5095134	6/10/2020	Injury	DEXCOM, INC.	6/19/2020	MDS	SENSORS	Patient-Device Incompatibility	Itching Sensation; Rash; Skin Irritation
MW5095056	6/15/2020	Injury	DEXCOM, INC.	6/17/2020	MDS	DEXCOM G6	Patient-Device Incompatibility	Irritation; Itching Sensation; Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5094953	6/8/2020	Malfunction	DEXCOM, INC.	6/11/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Break	No Known Impact Or Consequence To Patient
MW5094953	6/8/2020	Malfunction	DEXCOM, INC.	6/11/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Break	No Known Impact Or Consequence To Patient
MW5094889	6/4/2020	Injury	DEXCOM INC.	6/8/2020	QBJ	DEXCOM G6 CGM	Patient-Device Incompatibility	Pain; Skin Irritation; No Code Available
MW5094824	6/1/2020	Injury	DEXCOM, INC.	6/3/2020	QBJ	DEXCOM	Patient-Device Incompatibility	Rash
MW5094804	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Rash; Skin Irritation
MW5094804	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	MDS	DEXCOM SENSOR	Patient-Device Incompatibility	Rash; Skin Irritation
MW5094804	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	MDS	DEXCOM SENSOR	Patient-Device Incompatibility	Rash; Skin Irritation
MW5094804	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	MDS	DEXCOM SENSOR	Patient-Device Incompatibility	Rash; Skin Irritation
MW5094808	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Local Reaction; Skin Irritation; Blood Loss
MW5094808	5/29/2020	Injury	DEXCOM, INC.	6/2/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash; Local Reaction; Skin Irritation; Blood Loss

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50947 18	5/22/2020	Injury	DEXCOM, INC.	5/27/2020	MDS	DEXCOM CGM	Product Quality Problem	Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined
MW50947 18	5/22/2020	Injury	DEXCOM, INC.	5/27/2020	MDS	DEXCOM CGM	Product Quality Problem	Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined
MW50947 18	5/22/2020	Injury	DEXCOM, INC.	5/27/2020	MDS	DEXCOM CGM	Product Quality Problem	Erythema; Hemorrhage/Bleeding; Pain; Swelling; Complaint, Ill-Defined
MW50947 20	5/25/2020	Malfunction	DEXCOM, INC.	5/27/2020	QBJ	DEXCOM G6 SENSOR	Separation Failure	No Known Impact Or Consequence To Patient
MW50947 21	1/21/2020	Injury	DEXCOM, INC.	5/27/2020	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Itching Sensation; Rash
MW50947 28	5/19/2020	Injury	DEXCOM, INC.	5/27/2020	MDS	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Patient-Device Incompatibility	Irritation; Itching Sensation; Burning Sensation
MW50945 78	3/1/2020	Injury	DEXCOM, INC.	5/19/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5094539	5/1/2020	Injury	DEXCOM, INC.	5/18/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Impaired Healing; Caustic/Chemical Burns
MW5094500	5/2/2020	Injury	DEXCOM, INC.	5/14/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Itching Sensation; Rash
MW5094500	5/2/2020	Injury	DEXCOM, INC.	5/14/2020	MDS	DEXCOM G6 SENSOR	Adverse Event Without Identified Device or Use Problem	Hypersensitivity/Allergic reaction; Itching Sensation; Rash
MW5094502	2/20/2020	Injury	DEXCOM, INC.	5/14/2020	MDS	DEXCOM G6 SENSOR	Patient-Device Incompatibility	Itching Sensation; Pain; Rash; Scar Tissue; Discharge; Caustic/Chemical Burns
MW5094506	4/1/2020	Injury	DEXCOM, INC.	5/14/2020	QBJ	DEXCOM G6	Patient-Device Incompatibility	Reaction
MW5094464	5/8/2020	Injury	DEXCOM, INC.	5/12/2020	QDK	DEXCOM G6	Patient-Device Incompatibility	Rash; Blood Loss
MW5094465	4/6/2020	Injury	DEXCOM, INC.	5/12/2020	QBJ	DEXCOM G6 CGM	Patient-Device Incompatibility	Skin Irritation; Blood Loss
MW5094390	5/1/2020	Injury	DEXCOM, INC.	5/5/2020	QBJ	DEXCOM G6	Product Quality Problem	Erythema; Itching Sensation; Rash; Reaction

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50943 90	5/1/2020	Injury	DEXCOM, INC.	5/5/2020	QBJ	DEXCOM G6	Product Quality Problem	Erythema; Itching Sensation; Rash; Reaction
MW50943 90	5/1/2020	Injury	DEXCOM, INC.	5/5/2020	QBJ	DEXCOM G6	Product Quality Problem	Erythema; Itching Sensation; Rash; Reaction
MW50943 91	5/1/2020	Injury	DEXCOM, INC.	5/5/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR	Product Quality Problem	Rash; Swelling; Reaction
MW50942 35	4/19/2020	Injury	DEXCOM, INC.	4/22/2020	QBJ	DEXCOM G6 CONSTANT GLUCOSE MONITOR	Difficult to Insert; Device Dislodged or Dislocated; Output Problem	Pain
MW50939 25	3/24/2020	Malfunction	DEXCOM, INC.	3/25/2020	QBJ	DEXCOM G6 CGM SENSOR	Device Alarm System; Battery Problem	Hyperglycemia
MW50939 25	3/24/2020	Malfunction	DEXCOM, INC.	3/25/2020	QBJ	DEXCOM G6 CGM SENSOR	Device Alarm System; Battery Problem	Hyperglycemia
MW50939 25	3/24/2020	Malfunction	DEXCOM, INC.	3/25/2020	QBJ	DEXCOM G6 CGM SENSOR	Device Alarm System; Battery Problem	Hyperglycemia
MW50934 87	12/28/2019	Injury	DEXCOM, INC.	3/2/2020	QBJ	DEXCOM G6 CGM	Retraction Problem; Manufacturing, Packaging or Shipping Problem	Hemorrhage/Bleeding; Pain

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5093487	12/28/2019	Injury	DEXCOM, INC.	3/2/2020	QBJ	DEXCOM G6 CGM	Retraction Problem; Manufacturing, Packaging or Shipping Problem	Hemorrhage/Bleeding; Pain
MW5093247	2/12/2020	Injury	DEXCOM, INC.	2/24/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Device Fell	No Known Impact Or Consequence To Patient
MW5093055	2/13/2020	Injury	DEXCOM, INC.	2/18/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITORING DEVICE	Defective Device	No Known Impact Or Consequence To Patient
MW5092936	2/7/2020	Injury	DEXCOM, INC.	2/11/2020	QBJ	DEXCOM G6	Difficult to Remove	No Known Impact Or Consequence To Patient
MW5092767	1/11/2020	Malfunction	DEXCOM, INC.	2/4/2020	QBJ	DEXCOM G6	Connection Problem	No Known Impact Or Consequence To Patient
MW5092601	1/24/2020	Malfunction	DEXCOM, INC.	1/28/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem	No Known Impact Or Consequence To Patient

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5092601	1/24/2020	Malfunction	DEXCOM, INC.	1/28/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem	No Known Impact Or Consequence To Patient
MW5092601	1/24/2020	Malfunction	DEXCOM, INC.	1/28/2020	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Electrical /Electronic Property Problem; Communication or Transmission Problem; Device Sensing Problem	No Known Impact Or Consequence To Patient
MW5092555	1/24/2020	Injury	DEXCOM, INC.	1/27/2020	QBJ	DEXCOM G6	Mechanical Problem	No Known Impact Or Consequence To Patient
MW5092455	1/17/2020	Injury	DEXCOM, INC.	1/22/2020	QDK	DEXCOM G6CONTINUOUS GLUCOSE MONITORING DEVICE	Computer Software Problem; Loss of Data	Hypoglycemia
MW5092457	1/11/2020	Injury	DEXCOM, INC.	1/22/2020	QDK	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Retraction Problem; Physical Resistance/Sticking	Pain

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5092457	1/11/2020	Injury	DEXCOM, INC.	1/22/2020	QDK	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Retraction Problem; Physical Resistance/Sticking	Pain
MW5092457	1/11/2020	Injury	DEXCOM, INC.	1/22/2020	QDK	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Retraction Problem; Physical Resistance/Sticking	Pain
MW5091938	12/26/2019	Malfunction	DEXCOM, INC.	12/30/2019	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Unable to Obtain Readings	No Known Impact Or Consequence To Patient
MW5091396	11/30/2019	Injury	DEXCOM, INC.	12/3/2019	QBJ	DEXCOM G6	Disconnection; Incorrect, Inadequate or Imprecise Result or Readings; Low Test Results; Device Displays Incorrect Message	Hyperglycemia
MW5091397	12/1/2019	Injury	DEXCOM, INC.	12/3/2019	QBJ	DEXCOM G6	Wireless Communication Problem	No Known Impact Or Consequence To Patient
MW5091440	11/30/2019	Malfunction	DEXCOM, INC.	12/3/2019	QBJ	DEXCOM G6	Application Program Problem	Hypoglycemia; Cognitive Changes

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50905 42	3/10/2011	Injury	DEXCOM, INC.	10/21/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings	Hypoglycemia
MW50905 42	3/10/2011	Injury	DEXCOM, INC.	10/21/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings	Hypoglycemia
MW50905 42	3/10/2011	Injury	DEXCOM, INC.	10/21/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings	Hypoglycemia
MW50905 42	3/10/2011	Injury	DEXCOM, INC.	10/21/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings	Hypoglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5090542	3/10/2011	Injury	DEXCOM, INC.	10/21/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Defective Component; High Readings	Hypoglycemia
MW5090423	10/13/2019	Malfunction	DEXCOM, INC.	10/15/2019	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW5089729	9/9/2019	Malfunction	DEXCOM, INC.	9/11/2019	QBJ	DEXCOM TRANSMITTER	Electromagnetic Interference	No Known Impact Or Consequence To Patient
MW5089688	9/8/2019	Injury	DEXCOM, INC.	9/10/2019	QBJ	DEXCOM G6 SENSOR	Defective Component	No Known Impact Or Consequence To Patient
MW5089421	8/26/2019	Injury	DEXCOM, INC.	8/28/2019	QBJ	DEXCOM G6	High Readings	No Known Impact Or Consequence To Patient
MW5089191	8/15/2019	Malfunction	DEXCOM, INC.	8/19/2019	QBJ	DEXCOM G6 CGM	Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50887 68	10/29/2019	Malfunction	DEXCOM, INC.	8/5/2019	QBJ	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50887 68	10/29/2019	Malfunction	TANDEM DIABETES CARE, INC.	8/5/2019	PQF	T:SLIMX2	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50887 68	10/29/2019	Malfunction	DEXCOM INC.	8/5/2019	---	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50887 68	10/29/2019	Malfunction	DEXCOM, INC.	8/5/2019	---	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50887 68	10/29/2019	Malfunction	DEXCOM, INC.	8/5/2019	---	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50887 68	10/29/2019	Malfunction	DEXCOM, INC.	8/5/2019	---	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50887 68	10/29/2019	Malfunction	DEXCOM, INC.	8/5/2019	---	DEXCOM G6 CGM SENSOR	Device Difficult to Program or Calibrate; Incorrect, Inadequate or Imprecise Result or Readings	No Known Impact Or Consequence To Patient
MW50885 41	7/25/2019	Injury	DEXCOM, INC.	7/29/2019	QBJ	DEXCOM G6 SENSOR	Break; Sensing Intermittently	No Known Impact Or Consequence To Patient
MW50855 74	3/31/2019	Injury	DEXCOM, INC	4/3/2019	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Defective Device	Pain; Blood Loss

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5085574	3/31/2019	Injury	DEXCOM, INC	4/3/2019	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Defective Device	Pain; Blood Loss
MW5085574	3/31/2019	Injury	DEXCOM, INC	4/3/2019	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Defective Device	Pain; Blood Loss
MW5084988	3/14/2019	Malfunction	DEXCOM INC.	3/18/2019	QDK	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Detachment of Device or Device Component; Material Deformation	No Known Impact Or Consequence To Patient
MW5084487	2/18/2019	NA	DEXCOM INC.	2/27/2019	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Display or Visual Feedback Problem; Therapeutic or Diagnostic Output Failure	Hypoglycemia
MW5083130	1/10/2019	Injury	DEXCOM, INC.	1/15/2019	QBJ	DEXCOM G6	False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem	Hypoglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW50831 30	1/10/2019	Injury	DEXCOM, INC.	1/15/2019	QBJ	DEXCOM G6	False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem	Hypoglycemia
MW50831 30	1/10/2019	Injury	DEXCOM, INC.	1/15/2019	QBJ	DEXCOM G6	False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem	Hypoglycemia
MW50831 30	1/10/2019	Injury	DEXCOM, INC.	1/15/2019	QBJ	DEXCOM G6	False Alarm; Incorrect, Inadequate or Imprecise Result or Readings; Protective Measures Problem	Hypoglycemia
MW50829 36	1/4/2019	Malfunction	DEXCOM INC.	1/7/2019	QDK	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR	Material Integrity Problem	No Known Impact Or Consequence To Patient
MW51617 50	2024/10/19 4:00:00	Malfunction	DEXCOM, INC.	2024/10/28 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5160849	2024/06/20 4:00:00	Malfunction	DEXCOM, INC	2024/10/08 4:00:00	QBJ	DEXCOM G7		Device Embedded In Tissue or Plaque
MW5160431	2024/07/26 4:00:00	Malfunction	DEXCOM, INC.	2024/09/30 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5160432	2024/07/31 4:00:00	Malfunction	DEXCOM, INC.	2024/09/30 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5160433	2024/08/10 4:00:00	Malfunction	DEXCOM, INC.	2024/09/30 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5160434	2024/09/22 4:00:00	Malfunction	DEXCOM, INC.	2024/09/30 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159660	2024/09/14 4:00:00	Malfunction	DEXCOM, INC.	2024/09/17 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE SENSOR		Insufficient Information
MW5159661	2024/09/14 4:00:00	Malfunction	DEXCOM, INC.	2024/09/17 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE SENSOR		No Clinical Signs, Symptoms or Conditions
MW5159662	2024/09/14 4:00:00	Malfunction	DEXCOM, INC.	2024/09/17 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE SENSOR		No Clinical Signs, Symptoms or Conditions
MW5159553	2024/09/06 4:00:00	Injury	DEXCOM, INC.	2024/09/11 4:00:00	QBJ	DEXCOM G7		Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5159459	2024/08/23 4:00:00	Injury	DEXCOM, INC.	2024/09/09 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		Insufficient Information
MW5159232	2024/08/13 4:00:00	Malfunction	DEXCOM, INC.	2024/09/05 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159233	2024/08/13 4:00:00	Malfunction	DEXCOM, INC.	2024/09/05 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159234	2024/08/13 4:00:00	Malfunction	DEXCOM, INC.	2024/09/05 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159270	2024/09/03 4:00:00	Malfunction	DEXCOM, INC.	2024/09/05 4:00:00	QBJ	DEXCOM G7		Insufficient Information
MW5158720	2024/08/16 4:00:00	Injury	DEXCOM, INC.	2024/08/21 4:00:00	QBJ	DEXCOM G7		Insufficient Information
MW5158742	2024/08/09 4:00:00	Malfunction	DEXCOM, INC.	2024/08/21 4:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5158743	2024/08/09 4:00:00	Malfunction	DEXCOM, INC.	2024/08/21 4:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5158744	2024/08/09 4:00:00	Malfunction	DEXCOM, INC.	2024/08/21 4:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5157923	2024/07/27 4:00:00	Injury	DEXCOM, INC.	2024/08/01 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW5157752	2024/04/17 4:00:00	Injury	DEXCOM, INC.	2024/07/26 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW5157005	2024/06/30 4:00:00	Malfunction	DEXCOM INC	2024/07/03 4:00:00	QBJ	DEXCOM G7 SENSOR		Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5157006	2024/06/30 4:00:00	Malfunction	DEXCOM INC.	2024/07/03 4:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5157007	2024/06/30 4:00:00	Malfunction	DEXCOM INC.	2024/07/03 4:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5157008	2024/06/30 4:00:00	Malfunction	DEXCOM INC.	2024/07/03 4:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5157039	2024/06/29 4:00:00	Injury	DEXCOM, INC.	2024/07/03 4:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5156887	2024/06/27 4:00:00	Malfunction	DEXCOM, INC.	2024/07/01 4:00:00	QBJ	DEXCOM G7 SENSORS		Insufficient Information
MW5156888	2024/06/27 4:00:00	Malfunction	DEXCOM, INC.	2024/07/01 4:00:00	QBJ	DEXCOM G7 SENSORS		Insufficient Information
MW5156889	2024/06/27 4:00:00	Malfunction	DEXCOM, INC.	2024/07/01 4:00:00	QBJ	DEXCOM G7 SENSORS		Insufficient Information
MW5156912	2024/06/26 4:00:00	Malfunction	DEXCOM, INC.	2024/07/01 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW5156744	2024/06/18 4:00:00	Injury	DEXCOM, INC.	2024/06/27 4:00:00	QBJ	DEXCOM G7		Hemorrhage/Bleeding
MW5156745	2024/06/18 4:00:00	Injury	DEXCOM, INC.	2024/06/27 4:00:00	QBJ	DEXCOM G7		Hemorrhage/Bleeding
MW5156699	2024/06/03 4:00:00	Injury	DEXCOM, INC.	2024/06/26 4:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5156572	2024/06/21 4:00:00	Injury	DEXCOM, INC.	2024/06/25 4:00:00	QBJ	DEXCOM G7 CGM		Hyperglycemia
MW5156573	2024/06/21 4:00:00	Injury	DEXCOM, INC.	2024/06/25 4:00:00	QBJ	DEXCOM G7 CGM		Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5156574	2024/06/21 4:00:00	Injury	DEXCOM, INC.	2024/06/25 4:00:00	QBJ	DEXCOM G7 CGM		Hyperglycemia
MW5156575	2024/06/21 4:00:00	Injury	DEXCOM, INC.	2024/06/25 4:00:00	QBJ	DEXCOM G7 CGM		Hyperglycemia
MW5156576	2024/06/21 4:00:00	Injury	DEXCOM, INC.	2024/06/25 4:00:00	QBJ	DEXCOM G7 CGM		Hyperglycemia
MW5153370	2024/03/13 4:00:00	Injury	DEXCOM, INC.	2024/03/28 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITORING SYSTEM		No Clinical Signs, Symptoms or Conditions
MW5152822	2024/03/11 4:00:00	Malfunction	DEXCOM, INC.	2024/03/13 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW5152286	2024/02/23 5:00:00	Injury	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152287	2024/02/23 5:00:00	Malfunction	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152288	2024/02/23 5:00:00	Malfunction	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152289	2024/02/23 5:00:00	Malfunction	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152290	2024/02/23 5:00:00	Malfunction	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152291	2024/02/23 5:00:00	Malfunction	DEXCOM, INC.	2024/03/04 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5152177	2024/02/01 5:00:00	Injury	DEXCOM, INC.	2024/02/27 5:00:00	QBJ	DEXCOM G7 SENSOR		Rash; Blister

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5151824	2024/02/15 5:00:00	Injury	DEXCOM, INC.	2024/02/21 5:00:00	QBJ	DEXCOM G7		Hypoglycemia
MW5151008	2024/01/30 5:00:00	Malfunction	DEXCOM, INC.	2024/02/02 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5151009	2024/01/30 5:00:00	Malfunction	DEXCOM, INC.	2024/02/02 5:00:00	QBJ	DEXCOM G7 SENSOR		No Clinical Signs, Symptoms or Conditions
MW5150844	2024/01/13 5:00:00	Injury	DEXCOM, INC.	2024/01/26 5:00:00	QBJ	DEXCOM G7		Insufficient Information
MW5150787	2024/01/17 5:00:00	Injury	DEXCOM, INC.	2024/01/25 5:00:00	QBJ	DEXCOM G7		Itching Sensation; Skin Inflammation/ Irritation
MW5150679	2024/01/11 5:00:00	Injury	DEXCOM, INC.	2024/01/23 5:00:00	QBJ	DEXCOM		Hyperglycemia; Diabetic Ketoacidosis
MW5150080	2023/12/30 5:00:00	Malfunction	DEXCOM, INC.	2024/01/09 5:00:00	QBJ	DEXCOM G7		Sleep Dysfunction
MW5149773	2023/12/27 5:00:00	Injury	DEXCOM INC.	2024/01/03 5:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW5149365	2023/12/08 5:00:00	Injury	DEXCOM, INC.	2023/12/18 5:00:00	QBJ	DEXCOM G7		Bacterial Infection; Septic Shock
MW5149167	2023/12/07 5:00:00	Injury	DEXCOM, INC.	2023/12/12 5:00:00	QBJ	DEXCOM G7		Needle Stick/Puncture
MW5149168	2023/12/07 5:00:00	Injury	DEXCOM, INC.	2023/12/12 5:00:00	QBJ	DEXCOM G7		Needle Stick/Puncture
MW5147980	2023/10/18 4:00:00	Injury	DEXCOM, INC.	2023/11/08 5:00:00	QBJ	DEXCOM G7 CGM SYSTEM		Anxiety; Sleep Dysfunction
MW5146655	2023/09/29 4:00:00	Injury	DEXCOM, INC.	2023/10/05 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITOR		Foreign Body In Patient

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5146255	2023/09/23 4:00:00	Malfunction	DEXCOM, INC.	2023/09/27 4:00:00	QBJ	DEXCOM G7 SENSOR		Insufficient Information
MW5145972	2023/09/17 4:00:00	Malfunction	DEXCOM, INC.	2023/09/20 4:00:00	QBJ	DEXCOM G7 CONTINUOUS GLUCOSE MONITORING		No Clinical Signs, Symptoms or Conditions
MW5145611	2023/08/04 4:00:00	Malfunction	DEXCOM INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM		No Clinical Signs, Symptoms or Conditions
MW5145617	2023/08/22 4:00:00	Malfunction	DEXCOM, INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM	Incorrect, Inadequate or Imprecise Result or Readings; Defective Device	No Clinical Signs, Symptoms or Conditions
MW5145618	2023/08/30 4:00:00	Malfunction	DEXCOM, INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM	Incorrect, Inadequate or Imprecise Result or Readings; Calibration Problem; Device Sensing Problem; Data Problem	No Clinical Signs, Symptoms or Conditions
MW5145620	2023/08/15 4:00:00	Malfunction	DEXCOM, INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM	Low Readings; Device Sensing Problem; Data Problem	No Clinical Signs, Symptoms or Conditions
MW5145621	2023/08/30 4:00:00	Malfunction	DEXCOM INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM		No Clinical Signs, Symptoms or Conditions
MW5145622	2023/09/07 4:00:00	Malfunction	DEXCOM INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW51456 23	2023/09/07 4:00:00	Malfunction	DEXCOM INC.	2023/09/13 4:00:00	QBJ	DEXCOM G7 CGM		No Clinical Signs, Symptoms or Conditions
MW51448 29	2023/08/17 4:00:00	Injury	DEXCOM, INC.	2023/08/21 4:00:00	QBJ	DEXCOM G7		No Clinical Signs, Symptoms or Conditions
MW51187 42	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	QBJ	DEXCOM G7 READER	Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem	Hypoglycemia
MW51187 43	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem	Hypoglycemia
MW51187 44	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem	Hypoglycemia
MW51187 45	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	No Audible Alarm; Product Quality Problem; Activation Failure; Wireless Communication Problem	Hypoglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW51187 46	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem	Hypoglycemia
MW51187 47	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	No Audible Alarm; Product Quality Problem; Incorrect, Inadequate or Imprecise Result or Readings; Activation Failure; Wireless Communication Problem	Hypoglycemia
MW51187 48	2023/06/20 4:00:00	Injury	DEXCOM, INC.	2023/06/23 4:00:00	MDS	DEXCOM G7 SENSORS	Defective Alarm; Product Quality Problem; Device Sensing Problem; Wireless Communication Problem	Hypoglycemia
MW51176 16	2023/05/01 4:00:00	Malfunction	DEXCOM, INC.	2023/05/17 4:00:00	QBJ	DEXCOMG7	Use of Device Problem; Difficult to Open or Remove Packaging Material	No Clinical Signs, Symptoms or Conditions
MW51176 17	2023/05/01 4:00:00	Malfunction	DEXCOM, INC.	2023/05/17 4:00:00	QBJ	DEXCOMG7	Use of Device Problem	Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW51176 18	2023/05/01 4:00:00	Malfunction	DEXCOM, INC.	2023/05/17 4:00:00	QBJ	DEXCOMG7	Product Quality Problem; Defective Component	No Clinical Signs, Symptoms or Conditions
MW51171 42	2023/04/26 4:00:00	Injury	DEXCOM, INC.	2023/04/28 4:00:00	QBJ	DEXCOM G7	Use of Device Problem; Communication or Transmission Problem	Hypoglycemia
MW51171 43	2023/04/26 4:00:00	Injury	DEXCOM, INC.	2023/04/28 4:00:00	MDS	DEXCOM G7 SENSOR	Use of Device Problem; Communication or Transmission Problem	Hypoglycemia
MW51594 27	8/30/2024 4:00	Malfunction	DEXCOM, INC.	9/9/2024 4:00	SAF	DEXCOM STELO CGM		No Clinical Signs, Symptoms or Conditions
2954323- 2024- 43124	2024/10/30 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/31 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Low Readings	Loss of consciousness; Confusion/ Disorientation
2954323- 2024- 43147	2024/10/23 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/31 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness
2954323- 2024- 42804	2024/10/15 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/30 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323- 2024- 42665	2024/10/16 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/29 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-42685	2024/10/21 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/29 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-42774	2024/10/22 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/29 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness
2954323-2024-42470	2024/10/21 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/28 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-42410	2024/09/28 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/26 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Device Displays Incorrect Message	Hypoglycemia
2954323-2024-42101	2024/10/15 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/24 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41953	2024/10/09 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/23 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Diabetic Ketoacidosis
2954323-2024-41652	2024/09/23 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/22 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	High Readings	Hypoglycemia; Loss of consciousness
2954323-2024-41718	2024/10/18 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/22 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-41720	2024/10/02 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/22 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41726	2024/10/10 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/22 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41738	2024/10/09 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/22 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41397	2024/10/08 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/21 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41590	2024/10/18 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/21 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Loss of consciousness
2954323-2024-41326	2024/10/17 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/18 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings	Loss of consciousness
2954323-2024-40975	2024/10/11 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-40986	2024/10/03 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-41022	2024/10/11 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-41105	2024/10/14 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia
2954323-2024-41128	2024/10/08 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Low Readings	Loss of consciousness
2954323-2024-41131	2024/09/27 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Headache
2954323-2024-40443	2024/09/30 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/14 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-40379	2024/09/30 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/13 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Low Readings	Hyperglycemia
2954323-2024-40091	2024/10/06 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/09 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hyperglycemia
2954323-2024-39716	2024/09/13 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/08 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-39781	2024/09/11 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/08 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia
2954323-2024-39850	2024/09/12 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/08 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Detachment of Device or Device Component	Hyperglycemia
2954323-2024-39864	2024/09/11 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/08 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Device Displays Incorrect Message	Hypoglycemia
2954323-2024-39865	2024/09/10 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/08 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Device Displays Incorrect Message	Hypoglycemia
2954323-2024-39621	2024/09/09 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/07 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Low Readings	Hyperglycemia; Loss of consciousness; Diaphoresis
2954323-2024-39385	2024/09/11 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/04 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38962	2024/09/26 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/02 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38696	2024/09/17 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/01 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38725	2024/10/16 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/01 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-38356	2024/09/24 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/30 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38385	2024/09/14 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/30 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38071	2024/09/16 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/27 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-38177	2024/09/24 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/27 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-37819	2024/09/16 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/25 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-37836	2024/08/28 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/25 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Fatigue; Loss of consciousness
2954323-2024-37452	2024/09/12 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/24 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-37556	2024/07/28 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/24 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	High Readings; Low Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-37602	2024/08/27 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/24 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness
2954323-2024-37199	2024/09/11 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/23 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-37312	2024/09/16 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/23 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-37384	2024/09/09 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/23 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia
2954323-2024-37068	2024/09/06 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/20 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-36432	2024/08/27 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/18 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-36119	2024/08/29 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-36282	2024/08/08 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/17 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Detachment of Device or Device Component	Fatigue; Headache; Vomiting; Dizziness; Discomfort; Diabetic Ketoacidosis; Loss of consciousness; Cognitive Changes
2954323-2024-35875	2024/09/01 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/16 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia
2954323-2024-35555	2024/09/09 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/13 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-35464	2024/09/05 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/12 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-35298	2024/08/14 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/11 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Adverse Event Without Identified Device or Use Problem	Pain; Skin Infection
2954323-2024-35302	2024/08/29 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/11 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	High Readings	Hypoglycemia
2954323-2024-34458	2024/08/08 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/09/05 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
2954323-2024-34164	2024/08/26 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/04 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-33175	2024/08/21 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/08/29 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-32833	2024/08/16 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/08/28 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-32980	2024/08/12 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/08/28 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-32298	2024/08/15 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/08/26 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-31811	2024/06/25 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/08/23 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia; Confusion/ Disorientation
2954323-2024-31398	2024/08/10 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/08/21 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-29693	2024/07/12 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/08/11 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS SENSOR	Unable to Obtain Readings; Device Displays Incorrect Message	Insufficient Information
2954323-2024-29653	2024/07/12 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/08/09 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Detachment of Device or Device Component	Hyperglycemia; Loss of consciousness

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-25655	2024/07/13 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/07/17 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-24171	2024/06/22 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/07/08 4:00:00	QBJ	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness; Diaphoresis
2954323-2024-21061	2024/05/23 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/06/18 4:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Dizziness; Confusion/ Disorientation
2954323-2023-49376	2023/10/09 4:00:00	Injury	ABBOTT DIABETES CARE INC	2023/11/09 5:00:00	QLG	FREESTYLE LIBRE 2 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness; Diaphoresis
2954323-2024-42946	2024/10/07 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/30 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS	Adverse Event Without Identified Device or Use Problem	Erythema; Pain; Skin Infection
2954323-2024-42379	2024/10/17 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/25 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Loss of consciousness; Shaking/Tremors
2954323-2024-41256	2024/09/25 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/18 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
2954323-2024-40546	2024/09/28 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/10/15 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-40525	2024/09/24 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/10/14 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hyperglycemia; Dizziness
2954323-2024-38072	2024/09/02 4:00:00	Malfunction	ABBOTT DIABETES CARE INC	2024/09/27 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
2954323-2024-33024	2024/08/02 4:00:00	Injury	ABBOTT DIABETES CARE INC	2024/08/28 4:00:00	QBJ	FREESTYLE LIBRE 3 PLUS	Unable to Obtain Readings; Device Displays Incorrect Message	Hypoglycemia; Dizziness; Diaphoresis
MW5086 448	5/2/2019 0:00	Malfunction	MEDTRONIC , INC.	5/7/2019 0:00	MDS	MEDTRONIC GUARDIAN 3 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Low Test Results	Abdominal Pain; Hyperglycemia; Diabetic Ketoacidosis
30098627 00-2024-00899	2024/08/05 4:00:00	Injury	SENSEONICS INC.	2024/09/10 4:00:00	QHJ	EVERSENSE E3 SENSOR	Adverse Event Without Identified Device or Use Problem	Skin Infection; Skin Inflammation/ Irritation
Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
Column1	Column2	Column3	Column4	Column5	Column 6	Column7	Column8	Column9
MW51611 55	10/6/2024 4:00	Injury	DEXCOM, INC.	10/16/2024 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5160563		Death	DEXCOM, INC.	10/3/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		Insufficient Information
MW5160567	9/27/2024 4:00	Injury	DEXCOM, INC.	10/3/2024 4:00	QBJ	DEXCOM G6 SENSOR		Hyperglycemia; Hypoglycemia
MW5160467		Injury	DEXCOM, INC.	10/2/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions
MW5160472		Injury	DEXCOM, INC.	10/2/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5160333		Death	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5160384		Malfunction	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR SYSTEM		Rash
MW5160420		Injury	DEXCOM, INC.	9/30/2024 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5160191		Injury	DEXCOM, INC.	9/26/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		Hyperglycemia
MW5160099	9/19/2024 4:00	Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM G6 SENSOR		Insufficient Information
MW5160103		Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5160104		Malfunction	DEXCOM, INC.	9/25/2024 4:00	QBJ	DEXCOM		No Clinical Signs, Symptoms or Conditions
MW5160013	9/16/2024 4:00	Injury	DEXCOM, INC.	9/24/2024 4:00	QBJ	DEXCOM SENSOR		Insufficient Information
MW5160023	9/17/2024 4:00	Injury	DEXCOM, INC.	9/24/2024 4:00	QBJ	DEXCOM CONTINUOUS GLUCOSE MONITOR		No Clinical Signs, Symptoms or Conditions
MW5159708	9/16/2024 4:00	Injury	DEXCOM, INC.	9/17/2024 4:00	QBJ	DEXCOM BLOOD GLUCOSE MONITOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5159446	9/5/2024 4:00	Injury	DEXCOM, INC.	9/9/2024 4:00	QBJ	DEXCOM G6 SENSOR	High Test Results	Insufficient Information
MW5159455	8/2/2024 4:00	Injury	DEXCOM, INC.	9/9/2024 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5159311	8/31/2024 4:00	Injury	DEXCOM, INC.	9/6/2024 4:00	QBJ	DEXCOM SENSOR		No Clinical Signs, Symptoms or Conditions
MW5159349	9/4/2024 4:00	Injury	DEXCOM, INC.	9/6/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5159056	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5159057	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5159058	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5159081	8/26/2024 4:00	Injury	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6	High Readings; Device Sensing Problem	Hyperglycemia
MW5159137	8/27/2024 4:00	Malfunction	DEXCOM, INC.	8/29/2024 4:00	QBJ	DEXCOM G6 SENSOR		Insufficient Information
MW5158710		Injury	DEXCOM, INC.	8/21/2024 4:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5158575	8/14/2024 4:00	Injury	DEXCOM, INC.	8/16/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5158302		Injury	DEXCOM, INC.	8/9/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5158257	8/5/2024 4:00	Injury	DEXCOM, INC.	8/7/2024 4:00	QBJ	DEXCOM	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5157864		Injury	DEXCOM, INC.	7/30/2024 4:00	QBJ	DEXCOM		Hypoglycemia
MW5157737	7/22/2024 4:00	Injury	DEXCOM, INC.	7/26/2024 4:00	QBJ	DEXCOM G6 SENSOR		Discomfort
MW5156911	6/26/2024 4:00	Malfunction	DEXCOM, INC.	7/1/2024 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5156505	6/17/2024 4:00	Injury	DEXCOM, INC.	6/21/2024 4:00	QBJ	DEXCOM G6		Insufficient Information; No Clinical Signs, Symptoms or Conditions
MW5156459		Injury	DEXCOM, INC.	6/20/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5156471		Injury	DEXCOM, INC.	6/20/2024 4:00	QBJ	DEXCOM		Insufficient Information
MW5155988		Injury	DEXCOM, INC.	6/7/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5155768	6/3/2024 4:00	Malfunction	DEXCOM, INC.	6/4/2024 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5153878	4/11/2024 4:00	Malfunction	DEXCOM, INC.	4/15/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5153513	4/3/2024 4:00	Injury	DEXCOM, INC.	4/3/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia
MW5153215	3/23/2024 4:00	Injury	DEXCOM, INC.	3/26/2024 4:00	QBJ	DEXCOM SENSOR	High Readings	No Clinical Signs, Symptoms or Conditions
MW5153027	3/18/2024 4:00	Injury	DEXCOM, INC.	3/20/2024 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5153005	2/20/2024 5:00	Malfunction	DEXCOM, INC.	3/19/2024 4:00	QBJ	DEXCOM G6 CONTINUEO US GLUCOSE MONITOR SYSTEM		Foreign Body In Patient
MW5152444		Injury	DEXCOM, INC.	3/6/2024 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5151801	2/11/2024 5:00	Injury	DEXCOM, INC.	2/21/2024 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5151825		Injury	DEXCOM, INC.	2/21/2024 5:00	QBJ	DEXCOM	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5151167	2/5/2024 5:00	Injury	DEXCOM, INC.	2/7/2024 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5150660	1/20/2024 5:00	Injury	DEXCOM, INC.	1/23/2024 5:00	QBJ	DEXCOM G6 SENSOR	High Readings	No Clinical Signs, Symptoms or Conditions
MW5150681	1/21/2024 5:00	Malfunction	DEXCOM, INC.	1/23/2024 5:00	QBJ	DEXCOM G6 TRANSMITTE R		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5150490		Injury	DEXCOM, INC.	1/17/2024 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings	Insufficient Information
MW5150101		Injury	DEXCOM, INC.	1/9/2024 5:00	QBJ	DEXCOM G6	High Readings; Device Sensing Problem	Hyperglycemia
MW5149414		Injury	DEXCOM, INC.	12/19/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia
MW5148679		Injury	DEXCOM, INC.	12/1/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	No Clinical Signs, Symptoms or Conditions
MW5148559		Injury	DEXCOM, INC.	11/29/2023 5:00	QBJ	DEXCOM G6	Incorrect, Inadequate or Imprecise Result or Readings; Failure to Sense	Hyperglycemia
MW5148569	11/23/2023 5:00	Injury	DEXCOM, INC.	11/29/2023 5:00	QBJ	DEXCOM G6 CGM SENSOR		Unspecified Infection; Rash
MW5148448		Injury	DEXCOM, INC.	11/22/2023 5:00	QBJ	DEXCOM G6	High Readings; Calibration Problem	Insufficient Information
MW5148455	11/18/2023 5:00	Malfunction	DEXCOM, INC.	11/22/2023 5:00	QBJ	DEXCOM G6 CGM SENSOR		No Clinical Signs, Symptoms or Conditions
MW5148368		Injury	DEXCOM, INC.	11/21/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5148145		Injury	DEXCOM, INC.	11/14/2023 5:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5148148		Injury	DEXCOM, INC.	11/14/2023 5:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5148101		Injury	DEXCOM, INC.	11/13/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	No Clinical Signs, Symptoms or Conditions
MW5148022		Injury	DEXCOM, INC.	11/9/2023 5:00	QBJ	DEXCOM G6		Hyperglycemia
MW5147844		Injury	DEXCOM, INC.	11/6/2023 5:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings	Hypoglycemia
MW5147801		Injury	DEXCOM, INC.	11/3/2023 4:00	QBJ	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Device Sensing Problem	Hyperglycemia
MW5147758		Injury	DEXCOM, INC.	11/2/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5147654		Injury	DEXCOM, INC.	11/1/2023 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5147085	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR		Insufficient Information

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5147086	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR		Insufficient Information
MW5147087	10/12/2023 4:00	Malfunction	DEXCOM, INC.	10/18/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE SENSOR	Defective Device	Insufficient Information
MW5146930	10/8/2023 4:00	Injury	DEXCOM, INC.	10/12/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia; Hypoglycemia; Diabetic Ketoacidosis
MW5146931		Injury	DEXCOM, INC.	10/12/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146871		Injury	DEXCOM, INC.	10/11/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5146505		Injury	DEXCOM, INC.	10/3/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5146340		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146355		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146362		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5146370		Injury	DEXCOM, INC.	9/29/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5146223	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR		No Clinical Signs, Symptoms or Conditions
MW5146224	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR SENSOR		No Clinical Signs, Symptoms or Conditions
MW5146225	9/25/2023 4:00	Malfunction	DEXCOM, INC.	9/27/2023 4:00	QBJ	DEXCOM G6 CONTINUOUS GLUCOSE MONITOR TRANSMITTER		No Clinical Signs, Symptoms or Conditions
MW5146203		Injury	DEXCOM, INC.	9/26/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146144	9/20/2023 4:00	Injury	DEXCOM, INC.	9/25/2023 4:00	QBJ	DEXCOM G6		Insufficient Information
MW5146150		Injury	DEXCOM, INC.	9/25/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5146110		Injury	DEXCOM, INC.	9/22/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145980		Injury	DEXCOM, INC.	9/20/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145922		Injury	DEXCOM, INC.	9/19/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW5145630	9/9/2023 4:00	Injury	DEXCOM, INC.	9/13/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5145634	9/9/2023 4:00	Injury	DEXCOM, INC.	9/13/2023 4:00	QBJ	DEXCOM		Skin Inflammation/ Irritation
MW5145381		Injury	DEXCOM, INC.	9/7/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145386		Injury	DEXCOM, INC.	9/7/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5145275		Injury	DEXCOM, INC.	9/6/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW5144951		Injury	DEXCOM, INC.	8/24/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5144952		Injury	DEXCOM, INC.	8/24/2023 4:00	QBJ	DEXCOM G6		Hypoglycemia
MW5144652		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5144656		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW5144657		Injury	DEXCOM, INC.	8/15/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions

Report Number	Event Date	Event Type	Manufacturer	Date Received	Product Code	Brand Name	Device Problem	Patient Problem
MW51228 81	7/25/2023 4:00	Injury	DEXCOM, INC.	8/9/2023 4:00	QBJ	DEXCOM G6		Dehydration; Diarrhea; Fatigue; Hyperglycemia; Vomiting; Diabetic Ketoacidosis
MW51206 29		Injury	DEXCOM, INC.	8/4/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW51203 50		Injury	DEXCOM, INC.	7/28/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW51202 17		Injury	DEXCOM, INC.	7/26/2023 4:00	QBJ	DEXCOM G6		Hyperglycemia
MW51199 83		Injury	DEXCOM, INC.	7/21/2023 4:00	QBJ	DEXCOM G6		No Clinical Signs, Symptoms or Conditions
MW51192 29	7/5/2023 4:00	Injury	DEXCOM, INC.	7/7/2023 4:00	QBJ	DEXCOM G6 SENSOR	Pumping Stopped; Low Readings	Hyperglycemia
MW51188 23	6/21/2023 4:00	Injury	DEXCOM, INC.	6/23/2023 4:00	MDS	DEXCOM G6 SENSOR	Incorrect, Inadequate or Imprecise Result or Readings; Therapeutic or Diagnostic Output Failure	Hypoglycemia

Abbreviations.